

Evidence Essentials

Genium® / Genium® X3 Microprocessor Knee

| | Mobility need or deficit of the patient | Evidence for benefits of the Genium / X3 compared to C-Leg |
|-----------------|--|--|
| Mobility | Patient feels limited in activities of daily living (ADL) and overall mobility | <ul style="list-style-type: none"> - Significant improvement in total score of Physical Functional Performance Test (10 ADLs), and subdomains Upper Body Function, Upper Body Strength, Lower Body Strength and Balance, scores no longer different from able-bodied individuals (Highsmith et al., 2016B; Mileusnic et al, 2019) - Significantly improved perceived ease and safety of ADLs (Hahn et al., 2016; Kannenberg et al., 2013; Mileusnic et al, 2019) - Significant improvements AMP scores and step-activity derived functional level (Highsmith et al., 2016A and 2016B; Mileusnic et al, 2019) |
| Mobility | Patient has difficulty with long-distance ambulation | <ul style="list-style-type: none"> - Significant reduction of stance-phase braking forces (Bellmann et al., 2012; Schmalz et al., 2014) - Optimized swing control with constant knee swing flexion angle of 64° across walking speed (Bellmann et al., 2012; Schmalz et al., 2014; Mileusnic et al, 2019) - Best correction of kinematic and kinetic gait deviations and compensatory mechanisms compared to NMPK and C-Leg (Varrecchia et al., 2019) |
| Mobility | Patient has difficulty negotiating stairs and clearing bigger obstacles | <ul style="list-style-type: none"> - Patients rated stair ascent and descent and clearing bigger obstacles significantly easier (Hahn et al., 2016; Highsmith et al., 2014B; Kannenberg et al., 2013; Mileusnic et al, 2019) - Many patients improve quality of stair ascent to reciprocal (step-over-step) gait pattern that can also be used for stepping over bigger obstacles (Aldridge et al., 2014; Bellmann et al., 2012A and 2012B; Highsmith et al., 2014A and 2016A; Mileusnic et al, 2019; Schmalz et al., 2014) |

Evidence Essentials

Genium® / Genium® X3 Microprocessor Knee

| | | |
|------------------------------|--|--|
| Mobility | Patient has difficulty negotiating slopes and uneven terrain | <ul style="list-style-type: none"> - Significantly increased self-selected walking speed and quality of slope descent (Bell et al., 2016; Mileusnic et al, 2019) - Significantly improved gait symmetry and quality and unloading of the sound limb during slope descent (Bellmann et al., 2012; Highsmith et al., 2016A, Lura et al., 2017; Schmalz et al., 2014; Mileusnic et al, 2019) - Significantly improved knee swing flexion (toe clearance during slope descent) (Bellmann et al., 2012; Highsmith et al., 2016A, Lura et al., 2017; Mileusnic et al, 2019; Schmalz et al., 2014) - Patients rated slope and uneven terrain ambulation significantly easier and safer (Hahn et al., 2016; Highsmith et al., 2014B; Kannenberg et al., 2013; Mileusnic et al, 2019) |
| Mobility | Patient has difficulty standing still for longer periods of time, especially on slopes and hills | <ul style="list-style-type: none"> - Significantly increased weight-bearing on the prosthetic side while standing still on level ground and slopes (Bellmann et al., 2012; Highsmith et al., 2014B) - Patients rate ADLs that require standing significantly easier and safer (Hahn et al., 2016; Kannenberg et al., 2013; Mileusnic et al, 2019) |
| Musculo-skeletal pain | Patient suffers from joint and back pain due to gait asymmetry and excessive loading | <ul style="list-style-type: none"> - Significant improvement in gait symmetry and, thus, loading of the locomotor system (Bellmann et al, 2012A; Highsmith et al., 2016C; Lura et al.; 2017; Mileusnic et al, 2019; Schmalz et al., 2014) - Best correction of kinematic and kinetic gait deviations and compensatory mechanisms compared to NMPK and C-Leg (Varrecchia et al., 2019) |

Evidence Essentials

Genium[®] / Genium[®] X3 Microprocessor Knee

References

Aldridge Whitehead JM, Wolf EJ, Scoville CR, Wilken JM. Does a microprocessor-controlled knee affect stair ascent strategies in persons with transfemoral amputation? Clin Orthop Rel Res 2014; 472(10): 3093-3101.

<http://www.ncbi.nlm.nih.gov/pmc/articles/PMC4160488/>

Bell EM, Pruziner AL, Wilken JM, Wolf EJ. Performance of conventional and X2(r) prosthetic knees during slope descent. Clin Biomech (Bristol, Avon) 2016 Mar; 33: 26-31. doi: 10.1016/j.clinbiomech.2016.01.008. Epub 2016 Feb 2.

<https://www.ncbi.nlm.nih.gov/pubmed/26921583>

Bellmann M, Schmalz T, Ludwigs E, Blumentritt S. Immediate effects of a new microprocessor-controlled prosthetic knee joint: a comparative biomechanical evaluation. Arch Phys Med Rehabil 2012A; 93(3): 541-549. [http://www.archives-pmr.org/article/S0003-9993\(11\)00944-0/abstract](http://www.archives-pmr.org/article/S0003-9993(11)00944-0/abstract)

[http://www.archives-pmr.org/article/S0003-9993\(11\)00944-0/abstract](http://www.archives-pmr.org/article/S0003-9993(11)00944-0/abstract)

Bellmann M, Schmalz T, Ludwigs E, Blumentritt S. Stair ascent with an innovative microprocessor-controlled exoprosthetic knee joint. Biomed Tech 2012B; 57(6): 435-444. <http://www.degruyter.com/view/j/bmte.2012.57.issue-6/bmt-2011-0029/bmt-2011-0029.xml>

<http://www.degruyter.com/view/j/bmte.2012.57.issue-6/bmt-2011-0029/bmt-2011-0029.xml>

Hahn A, Lang M, Stuckert C. Analysis of clinically important factors on the performance of advanced hydraulic, microprocessor-controlled exo-prosthetic knee joints based on 899 trial fittings. Medicine (Baltimore) 2016;95(45):e5386.

<https://www.ncbi.nlm.nih.gov/pmc/articles/PMC5106077/>

Highsmith MJ, Kahle JT, Lura DJ, Lewandowski AJ, Quillen WS, Kim HS. Stair ascent and ramp gait training with the Genium knee. Technol Innov 2014A; 15: 349-258.

<http://www.ingentaconnect.com/content/cog/ti/2014/00000015/00000004/art00011>

Highsmith MJ, Kahle JT, Miro RM, Lura DJ, Dubey RV, Carey SL, Quillen WS, Mengelkoch LJ. Perceived differences between the Genium und the C-leg microprocessor prosthetic knees in prosthetic-related function and quality of life. Technol Innov 2014B; 15: 269-375. <http://www.ingentaconnect.com/content/cog/ti/2014/00000015/00000004/art00013>

<http://www.ingentaconnect.com/content/cog/ti/2014/00000015/00000004/art00013>

Highsmith MJ, Klenow TD, Kahle JT, Wernke MM, Carey SL, Miro RM, Lura DJ, Sutton BS. Effects of the Genium knee system on functional level, stair ambulation, perceptive and economic outcomes in transfemoral amputees. Technol Innov 2016A; 18: 139-150. <http://dx.doi.org/10.21300/18.2-3.2016.139>.

<http://dx.doi.org/10.21300/18.2-3.2016.139>.

Highsmith MJ, Kahle JT, Miro RM, Cress EM, Lura DJ, Quillen WS, Carey SL, Dubey RV, Mengelkoch LJ. Functional performance differences between Genium and C-Leg prosthetic knees and intact knees. J Rehabil Res Dev 2016B;53(6):753-766. <http://dx.doi.org/10.1682/JRRD.2014.06.0149>.

<http://dx.doi.org/10.1682/JRRD.2014.06.0149>.

Highsmith MJ, Klenow TD, Kahle JT, Wernke MM, Carey SL, Miro RM, Lura DJ. Effects of the Genium microprocessor knee system on knee moment symmetry during hill walking. Technol Innov 2016C;18: 151-157.

<http://dx.doi.org/10.21300/18.2-3.2016.151>.

Kannenberg A, Zacharias B, Mileusnic M, Seyr M. Activities of daily living: Genium Bionic Prosthetic Knee compared with C-Leg. J Prosthet Orthot 2013; 25(3): 110-117.

http://journals.lww.com/jpojournal/Abstract/2013/07000/Activities_of_Daily_Living_Genium_Bionic.3.aspx

Evidence Essentials

Genium® / Genium® X3 Microprocessor Knee

Lura DJ, Wernke MM, Carey SL, Kahle JT, Miro RM, Highsmith MJ. Differences in knee flexion between the Genium and C-Leg microprocessor knees while walking on level ground and ramps. Clin Biomech (Bristol, Avon). 2015 Feb;30(2):175-81. doi: 10.1016/j.clinbiomech.2014.12.003. Epub 2014 Dec 13.

<http://www.sciencedirect.com/science/article/pii/S0268003314002988>

Mileusnic MP, Rettinger L, Highsmith MJ, Hahn A. Benefits of the Genium microprocessor controlled prosthetic knee on ambulation, mobility, activities of daily living and quality of life: a systematic literature review. Disabil Rehabil 2019 Aug 30:1-12. doi: 10.1080/17483107.2019.1648570. Online ahead of print.

Schmalz T, Bellmann M, Proebsting E, Blumentritt S. Effects of Adaptation to a Functionally New Prosthetic Lower-Limb Component: Results of Biomechanical Tests Immediately after Fitting and after 3 Months of Use. J Prosthet Orthot 2014; 26(3): 134-143.

http://journals.lww.com/jpojournal/Fulltext/2014/07000/Effects_of_Adaptation_to_a_Functionally_New.4.aspx

Varrecchia T, Serrao M, Rinaldi M, Ranavolo A, Conforto S, De Marchis C, Simonetti A, Poni I, Castellano S, Silveti A, Tatarelli A, Fiori L, Conte C, Draicchio F. Common and specific gait patterns in people with varying anatomical levels of lower-limb amputation and different prosthetic components. Hum Mov Sci 2019;66:9-21.

<https://doi.org/10.1016/j.humov.2019.03.008>