

Kenevo®

Reimbursement Guide

June 1, 2021



Quality for life

Information for practitioners and payers

Kenevo Reimbursement Guide

Product Information

June 2021

Indications

- (K2) Limited Community Ambulators
- (Low K3) Full Community Ambulators who walk with speeds of up to 1.9 mph and up to 1/2 mile outside the house per day.

¹Kenevo Coding (U.S. only)

The Healthcare Common Procedure Coding System (HCPCS) for prosthetics is an add-on code system. Rather than issuing new HCPCS Level II national codes to describe the various microprocessor knees that came to market, the Alpha-Numeric HCPCS Panel instead issued add-on codes (L5856, L5857, L5858 and L5859) to upgrade the mechanical (non-microprocessor) knee codes.

The following codes are PDAC verified for Kenevo, effective 12/24/2020:

L5828	Hydraulic Swing and Stance Phase Knee (base mechanical knee code)
L5845	Stance flexion feature
L5848	Stance extension damping feature
L5856	Microprocessor control feature, swing and stance phase, includes sensors

Warranty

Kenevo comes with a three-year manufacturer warranty (extendable to six years) which includes a complimentary condition-based service inspection within the 3-year term. During the warranty period, repair costs are covered except for those associated with damages resulting from improper use.

Health Canada Compliance

This device meets the requirements of the Medical Device Regulations (SOR/98-282). It has been classified as a class I medical device according to the classification criteria outlined in schedule 1 of the Medical Device Regulations.

FDA Status

Under FDA's regulations, the Kenevo Microprocessor-Controlled Prosthetic Knee is a Class I device, exempt from the premarket notification [510(k)] requirements. The Kenevo prosthetic knee has met all applicable control requirements which include Establishment Registration (21CFR 807), Medical Device Listing (21 CFR part 807), Quality System Regulation (21CFR part820), Labeling (21CFR part 801), and Medical Device Reporting (21 CFR Part 803). The Kenevo prosthetic knee is listed under JOINT, KNEE, EXTERNAL LIMB COMPONENT; Listing Number is E253231 and Manufacturer Registration Number is 3005190268.

Who Can Provide a Kenevo?

The Kenevo is prescribed by a physician and may only be provided by a qualified Prosthetist who has received specific product training. Ottobock employs a team of orthotists and prosthetists to educate practitioners on fabricating and fitting our products. This includes in-person and online training, webinars, and technical bulletins. We also provide Cooperative Care Services for the more challenging fittings, which includes on-site assistance with the fitting in conjunction with product qualification training for the practitioner.

¹ The product/device "Supplier" (defined as an O&P practitioner, O&P patient care facility, or DME supplier) assumes full responsibility for accurate billing of Ottobock products. It is the Supplier's responsibility to determine medical necessity; ensure coverage criteria is met; and submit appropriate HCPCS codes, modifiers, and charges for services/products delivered. It is also recommended that Supplier's contact insurance payer(s) for coding and coverage guidance prior to submitting claims. Ottobock Coding Suggestions and Reimbursement Guides are based on reasonable judgment and are not recommended to replace the Supplier's judgment. These recommendations may be subject to revision based on additional information or alpha-numeric system changes.

Kenevo Features and Benefits

Microprocessor Swing and Stance Phase Control (L5856)

The single-axis hydraulic unit of the Kenevo microprocessor knee is controlled by a microprocessor that continuously adjusts two independent servo motors, allowing for independent control of the valves controlling flexion and extension resistances during both stance and swing phase.

Input signals for the control algorithms of the Kenevo microprocessor are knee angle, axial load, and the three-dimensional movement of the knee in space detected by a knee angle sensor, an axial load sensor, and an inertial motion unit (IMU), which consist of accelerometers, gyroscopes and its own microprocessor, with all sensors located within the prosthetic knee. All sensor signals are sampled at a frequency of 100 Hz, allowing the microprocessor to precisely determine the exact phase of the gait cycle, stumbles, and the terrain the user is walking on. As a result, the microprocessor continuously and separately adjusts the necessary flexion and extension resistances of the hydraulic unit to safely and almost physiologically control stance and swing resistances in all gait situations [2, 3].

Stumble Recovery Plus

The microprocessor swing control of the Kenevo provides an enhanced stumble recovery feature. Stumble Recovery Plus allows for even higher weight-bearing capacity than adjusted for a regular stance phase and unique protection during swing knee flexion (heel rise). Like the C-Leg microprocessor knee, the hydraulic of the Kenevo employs two motorized valves to control resistance for knee flexion and extension separately. That allows for easy swing forward to ambulate while making sure, at the same time, that the knee can be loaded with the full body weight to recover from a stumble and prevent a fall. In addition to the Stumble Recovery feature of the C-Leg, the Kenevo further ramps-up knee flexion resistance (loading capacity) above the level for a regular stance phase in case of a disruption of gait during the entire swing phase including the phase of heel rise. In addition, swing release is based on the loading profile of each step and activated later than in microprocessor knees designed for higher-functioning individuals. This is because limited community ambulators usually walk slower, with more irregular gait (bigger variations from step-to-step), shuffling steps, or with additional walking aids such as crutches or a walker that result in reduced loading of the prosthesis. This function delivers much needed increased stability during late stance but makes sure that swing is released in every step to provide sufficient toe clearance. MPKs for higher-functioning individuals often

with irregular gait patterns. Also, at terminal swing, the microprocessor controls the dampening of the knee extension movement, preventing a hard impact that might impair the balance of the user.

Supported Safe Stand-to-Sit Function

The Kenevo automatically detects when your patient begins to sit down, adjusting the hydraulic resistance so the knee joint provides progressive support during sitting. This allows the amputee to shift the body weight to both legs and complete the sit down motion in a smooth, controlled manner at a controlled rate.

Once the amputee is seated, if the knee is still extended, Kenevo will relax into a seated position and will switch to energy-saving mode.

Benefits include:

- Supports safety and balance during sitting down
- Automatic unlock allows for hands-free operation without the need to unload the prosthesis, which is especially important for those who use walking aids such as canes or walkers.
- Relieves the contralateral side and increases the area of support by shifting load to both legs.



Supported Safe Sit-to-Stand Function

The Kenevo also automatically detects when the patient begins to stand up.

If the patient pauses during the standing-up motion, the knee will not collapse as long as the patient has made it at least half way (prosthesis has reached at least 45° flexion), which allows the patient to rest on the prosthesis and reposition their weight to the sound limb if standing up in a single motion is too tiring.

The knee switches automatically to Supported Stand-to-sit function if the patient tends to fall backward.

Inertial Motion Unit (IMU) Control

Backward Walking

This patented microprocessor control technology provides safety and stability when your patient is forced to step backward (e.g. when opening a door). Many microprocessor knees do not accommodate backward stepping, which may cause the knee to collapse if a backward step is taken.

Automatic Ramp Detection

The IMU on the Kenevo provides automatic ramp detection and provides additional assistance during ramp descent.

Intuitive Standing

Maintaining safety and balance while standing is critical for K2 and low-K3 patients. Kenevo allows the patient to intuitively stand on a flexed and stable knee when on level, uneven, or inclined surfaces (e.g. ramps and hills).

Contrast this to traditional K2 prosthetic knees, which require the user to extend the hip to stabilize the knee or cognitively ensure that their center of mass stays ahead of their knee axis to prevent unexpected buckling of the prosthetic knee.

Unlike mechanical knees, Kenevo offers clinicians a range of programmable stance stability options that can be customized to support each patient's individual capabilities.



Hydraulic Swing and Stance (L5828)

Hydraulic Stance Control

Hydraulic stance control provides resistance against knee flexion to support the body weight of the patient and prevent knee collapse. Knee flexion during weight-bearing is damped and controlled, mimicking the eccentric contraction of the quadriceps muscle during gait, to provide for shock absorption during level walking to

minimize hip and low-back stress. It also enables step-over-step slope and stair descent and uneven terrain ambulation, allowing patients to “ride” the knee when descending stairs and slopes.

Hydraulic Swing Control

The hydraulic swing phase control of Kenevo accommodates walking speeds of up to 3 km/h (1.9 mph). It also provides for terminal swing extension damping that prevents a hard terminal impact that would be hard to control for the patient by decelerating the prosthetic shank prior to heel strike. This mimics the eccentric contraction of the anatomical hamstrings and gluteus maximus muscle. Full extension is then reached smoothly in preparation for heel strike.

Hydraulic Stance Flexion (L5845)

Knee flexion during stance, i.e. during weight bearing, is important for level-walking as well as for the negotiation of uneven terrain, slopes and stairs. Non-amputated subjects control knee stance flexion with their muscles, specifically with the quadriceps muscle, and walk with a knee stance flexion of 15-25° on level ground. Individuals with an above-knee amputation can be fit with a prosthetic knee joint that allows for stance flexion during loading to improve shock absorption and relief of the hip and lumbar spine. The Kenevo has been proven to enable the patient to walk with stance flexion on level surfaces, uneven terrain, and obstacle courses. [2] Furthermore, Kenevo users can descend slopes and stairs with reciprocal (step-over-step) gait. All these activities require knee stance flexion during weight bearing.

Hydraulic Stance Extension Damping (L5848)

After the knee is flexed during stance phase (stance flexion), it needs to extend again to advance the body forward through mid-stance. This feature provides a smooth extension of knee. Without this function, the patient would feel a pronounced “snap back” or “jerk” at the knee, and would also present with an unnatural looking gait pattern. Energy is conserved by having this feature, as the patient will not have to attempt to control this motion with his residual limb muscles.

Knee Extension Assist

The knee extension assist is used in promoting knee extension at the beginning of swing phase extension. This function allows the user to walk more efficiently at variable cadence, since the spring extension assist mechanically limits the knee flexion at the end range and begins to bring the knee into extension for a more symmetrical gait at faster walking speeds. It also ensures the knee comes to full extension for the beginning of stance phase for a more secure loading condition during level walking, but in particular when descending stairs where full extension facilitates the positioning of the foot on the edge of a stair.

Intuitive Stationary Indoor Bicycle Function

The intuitive stationary indoor bicycle function is activated only when the user is seated. Once the user sits down and pedals, the Kenevo automatically switches to lower resistance after a few revolutions. The joint switches back to default settings for walking immediately when the user gets off the indoor bicycle. This mode can also be activated manually by the user via the Cockpit app.



Indications and contraindications

Ottobock recommends the following indications and contraindications. The practitioner should be familiar with these points when considering the Kenevo for each individual patient.

The Kenevo is appropriate for patients who

- Can walk at speeds of up to 1.9 mph
- Need a high level of safety while walking and standing
- Require a high degree of support while sitting down and standing up

Indications

- Amputation levels: knee disarticulation (unilateral & bilateral), transfemoral amputation (unilateral & bilateral), hip disarticulation
- K2, low-K3 functional levels*
- Body weight < 275 lbs.
- Patients who are physically and cognitively able to perceive visual / acoustic signals and /or mechanical vibrations

Contraindications

- Amputees with a K3 functional level who walk with speeds of 2 mph or greater and more than 1/2 mile outside the house per day.
- Amputees with K4 functional level

*More specifically, for patients who walk with speeds of up to 1.9 mph and up to 1/2 mile outside the house per day.

Evidence Summary

Kenevo[®] Microprocessor Stance Control Knee

	Mobility need or deficit of the patient	Proven benefits of MPK/Kenevo in K2 patients
Safety	Patient often stumbles and/or falls	MPK have been demonstrated to significantly reduce falls by up to 80%, significantly improve indicators for the risk of falling, and to reduce the frequency of stumbles.
Safety	Patient avoids activities of daily living due to safety concerns and lack of balance and/or balance confidence	MPK/Kenevo have been shown to improve the risk of falling, balance and balance confidence. This may result in the patient doing more activities with the prosthesis.
Slope negotiation	Patient has to ambulate on slopes/hills on a regular basis and struggles with slope descent and/or has to descend slopes and hills faster.	MPK have been shown to improve slope and hill negotiation with a more natural gait pattern and to significantly improve downhill walking speed.
Stair negotiation	Patient has to ambulate on stairs on a regular basis and struggles with stair descent, needs to descend stairs faster.	MPK have been demonstrated to significantly improve the quality of stair descent. This is an indicator of improved balance confidence and allows for descending stairs much faster.
Negotiation of uneven terrain /obstacles in the walkway	Patient has to ambulate on uneven terrain and/or clear obstacles in the walkway on a regular basis and struggles to do so and/or has to ambulate faster (e.g. for chasing kids).	MPK have been shown to have superior safety and allows for significantly walking faster on uneven terrain and obstacle courses.
Cognitive demand/multi-tasking during walking	Patient has to do concurrent activities while walking with the prosthesis on a regular basis and struggles with these activities (e.g. needs to stop walking or walk slower)	MPK have been demonstrated to increase multitasking capacities and cognitive burden while walking with the prosthesis.
Overall mobility	Patient is a limited community ambulator (MFCL-2, K2)	MPK have been shown to reduce uncontrolled falls by up to 80%; improve validated indicators of the risk of falling; increase walking speed on level ground by 14-25%, on uneven terrain by up to 20%, and slope descent by 30%; and improve stair negotiation. About 50% of K2 patients in the studies were able to improve their mobility to K3. Kenevo has been found to reduce additional wheelchair use from 87% to 37%.

Evidence Summary

Kenevo[®] Microprocessor Stance Control Knee

Safety: Reduced stumbles and falls

Several clinical and biomechanical studies have investigated the safety of prosthesis use as well as balance and balance confidence of individuals with patients with Medicare Functional Classification Level 2 (MFCL-2, K2, limited community ambulator) while walking with a prosthesis. A systematic review (1) analysed a total of six studies of sufficient methodological quality that compared the safety of microprocessor controlled prosthetic knees (MPK) with that of non-MP controlled prosthetic knees (non-MPK) in subjects with MFCL-2 mobility. Hafner et al. (2) and Kahle et al. (3) observed persons with a transfemoral amputation transitioning from a non-MPK to a C-Leg prosthesis to collect data on stumbles and falls. Hafner et al. (2) found a significant 80% reduction in the number of uncontrolled falls ($p < .01$) and a significant reduction in the frequency of stumbles ($p < .05$). Kahle et al. (3) reported a statistically significant 57% reduction in stumbles ($p = 0.006$) and a significant 64% reduction in falls ($p = .03$) in their mixed sample of subjects with MFCL-2 and -3 mobility. A systematic review that analysed only the subgroup of this study with MFCL-2 mobility found a significant 80% reduction in falls ($p < .05$) in this patient subgroup (1). The significant reduction in falls was recently confirmed by a study of Kaufman et al. (4) that enrolled 50 patients with MFCL-2 mobility who were randomized to transition from their customary non-MPK to one out of four different MPK ($p = .01$). Burnfield et al. (5) and Lansade et al. (6) studied the effect of using MP stance control knees, the Ottobock Compact (5) or Kenevo (6), respectively, on validated indicators of the risk of falling in individuals with MFCL-2 mobility. Compared to non-MPKs, the use of the Compact significantly improved the average time to complete the Timed-up-and-go-test (TUG) by 28% from 24.5 sec to 17.7 sec ($p = .018$) (5). In the study with 27 subjects using the Kenevo, the median time required to complete the TUG was significantly reduced from 21.4 sec to 17.9 sec ($p = .001$) (6). Thus, in both studies, the TUG time decreased well below the established threshold of 19 sec that indicates an increased risk of multiple falls in below-knee amputees when using a MP stance control knee (7).

Safety: Improved balance and balance confidence

Balance and balance confidence with the prosthesis are related to and/or associated with falling, fear of falling, and activity avoidance in persons with an above-knee amputation (8-14). Burnfield et al. assessed balance confidence in subjects with MFCL-2 mobility with the validated Activity-specific balance confidence (ABC) scale that improved significantly from 60.1 to 75.7 ($p = .001$) when using the MP stance control knee Compact as compared to non-MPKs (5). Scores below 67 indicate an increased risk of falling (8-10) and are associated with fear of falling and avoidance of activities (8, 15, 16). Kaufman et al., in their study that had enrolled 50 subjects with MFCL-2 mobility, saw a significant increase in median activity per day ($p = .02$) and spent significantly less time sitting ($p = .01$) when using a MPK, indicating an improved confidence in the prosthesis (4). This is supported by the findings of Theeven et al. (17) who found significantly improved performance in activities of daily living (ADL) and community ambulation in their study with 30 individuals with MFCL-2 mobility when using a MPK compared to their customary non-MPK. The biggest improvements were seen in activities that were performed with the arms and hands but required good lower body stability to execute (17). Likewise, Lansade et al. demonstrated a significant improvement in ADL performance assessed with the Locomotor Capabilities Index (LCI-5) in their study with 27 patients with MFCL-2 mobility using Kenevo as compared to

Evidence Summary

Kenevo[®] Microprocessor Stance Control Knee

non-MPKs (6). Similarly, Hafner et al. found a significantly increased multi-tasking ability ($p=.04$) of individuals with MFCL-2 mobility while walking with the C-Leg, indicating an improved confidence in the prosthesis (2). An observational study with 29 subjects assessing Kenevo as compared to NMPK typically prescribed and fit in patients with MFCL-2 mobility found that the percentage of participants who experienced no falls at all within 8 weeks increased from 45% to 72% (n.s.), the percentage of subjects who experienced no stumbles at all increased from 8% to 50% ($p=.044$), and 50% of individuals reported a reduction in the fear of falling when using the Kenevo (18).

Summarizing all studies with individuals with MFCL-2 mobility that had investigated the safety of MPKs compared to non-MPKs, a systematic review of the literature (1) and two more recent bigger clinical studies (4, 6) have concluded that MPK significantly reduced falls and significantly improved indicators for the risk of falling as well as balance and balance confidence.

References

1. Kannenberg A, Zacharias B, Pröbsting E: Benefits of microprocessor prosthetic knees to limited community ambulators: A systematic review. *J Rehabil Res Dev* 2014; 51 (10): 1469-1495.
2. Hafner BJ, Smith DG: Differences in function and safety between Medicare Functional Classification Level-2 and -3 transfemoral amputees and influence of prosthetic knee joint control. *J Rehabil Res Dev* 2009; 46 (3): 417-434.
3. Kahle JT, Highsmith MJ, Hubbard SL: Comparison of Non-microprocessor Knee Mechanism versus C-Leg on Prosthesis Evaluation Questionnaire, Stumbles, Falls, Walking Tests, Stair Descent, and Knee Preference; *J Rehabil Res Dev* 2008; 45 (1): 1-14.
4. Kaufman KR, Bernhardt KA, Symms K. Functional assessment and satisfaction of transfemoral amputees with mobility (FASTK2): A clinical trial of microprocessor-controlled vs. non-microprocessor-controlled knees. *Clin Biomech (Bristol, Avon)* 2018 Oct;58:116-122. doi: 10.1016/j.clinbiomech.2018.07.012. Epub 2018 Jul 19.
5. Burnfield JM, Eberly VJ, Gronely JK, Perry J, Yule WJ, Mulroy SJ. Impact of stance phase microprocessor-controlled knee prosthesis on ramp negotiation and community walking function in K2 level transfemoral amputees. *Prosthet Orthot Int* 2012, 36 (1): 95-104.
6. Lansade C, Vicaut E, Paysant J, Ménager D, Cristina MC, Braatz F, Domayer S, Pérennou D, Chiesa G. Mobility and safety with a microprocessor-controlled knee in moderately active amputees: A multi-centric randomized crossover trial. *Ann Phys Rehabil Med*, 2018 May 14. pii: S1877-0657(18) 30058-7. doi: 10.1016/j.rehab.2018.04.003. [Epub ahead of print]
7. Dite W, Connor HJ, Curtis HC. Clinical identification of multiple fall risk early after unilateral transtibial amputation. *Arch Phys Med Rehabil* 2007;88(1):109-114. PMID: 17207685.
8. Miller WC, Speechley M, Deathe AB. The prevalence and risk factors of falling and fear of falling among lower extremity amputees. *Arch Phys Med Rehabil* 2001;82(8):1031-1037. PMID: 11494181.
9. Gauthier-Gagnon C, Grise M-C, Potvin D. Predisposing factors related to prosthetic use by people with a transtibial and transfemoral amputation. *J Prosthet Orthotics* 1998; 10(4):99-109.
10. Miller WC, Deathe AB, Speechley M, Koval J: The influence of falling, fear of falling, and balance confidence on prosthetic mobility and social activity among individuals with a lower extremity amputation. *Arch Phys Med Rehabil* 2001; 82(9):1238-44.
11. Bertera EM, Bertera RL: Fear of falling and activity avoidance in a national sample of older adults in the United States. *Health Soc Work* 2008; 33:54-62.
12. Fletcher PC, Hirdes JP: Restriction in activity associated with fear of falling among community based seniors using home care services. *Age Ageing* 2004; 33: 273-279.
13. Delbaere K, Crombez G, Vanderstraeten G, Willems T, Cambier D: Fear-related avoidance of activities, falls and frailty. A prospective community based cohort study. *Age Ageing* 2004; 33: 368-373.

Evidence Summary

Kenevo[®] Microprocessor Stance Control Knee

14. Kempen GIJM, van Haastregt JCM, McKee KJ, Delbaere K, Zijlstra GAR: Socio-demographic, health-related and psychosocial correlates of fear of falling and avoidance of activity in community-living older persons who avoid activity due to fear of falling. *BMC Public Health* 2009; 9: 170-176.
15. Lajoie Y, Gallagher SP. Predicting falls within the elderly community: comparison of postural sway, reaction time, the Berg balance scale and the Activities-specific Balance Confidence (ABC) scale for comparing fallers and non-fallers. *Arch Gerontol Geriatr* 2004;38(1):11-26. PMID: 14599700.
16. Miller WC, Deathe AB, Speechley M. Psychometric properties of the Activities-specific Balance Confidence Scale among individuals with a lower-limb amputation. *Arch Phys Med Rehabil* 2003;84(5):656-661.
17. Theeven P, Hemmen B, Rings F, Meys G, Brink P, Smeets R, Seelen H. Functional added value of microprocessor controlled knee joints in daily life performance of Medicare Functional Classification Level-2 amputees. *J Rehabil Med.* 2011;43(10):906–15. [PMID:21947182]
18. Mileusnic M, Hahn A, Reiter S. Effects of a novel microprocessor-controlled knee, Kenevo, on the safety, mobility, and satisfaction of lower-activity patients with transfemoral amputation. *J Prosthet Orthot* 2017; 29(4): 198-205.

Improved slope descent

Ambulation on sloped terrain such as ramps and hills is associated with increased potential for slipping, loss of balance, and falling. Among the many causes for this is the fact that ramp and slope walking requires changes in the range of motion and strength compared to traditional stepping patterns used to traverse flat ground (2, 6). Able-bodied people use reciprocal (step-over-step) slope and hill descent, in which the supporting leg lowers the whole body down using knee flexion controlled by an eccentric contraction of the quadriceps muscle while the swinging leg swings and lands past the supporting leg. Usually, the step length is even between both legs. In above-knee amputees, most non-microprocessor controlled prosthetic knee mechanisms do not allow for any or enough knee flexion during weight-bearing to lower the body with the supporting prosthetic leg or are too difficult to control for most patients to do so safely (2, 7). That’s why above-knee amputees usually use a step-to or even a side-step pattern to descend slopes and hills. In the step-to pattern, the supporting sound leg lowers the body using knee flexion controlled by an eccentric contraction of the quadriceps while the prosthetic leg swings and lands past the sound leg. Then the sound leg is positioned next to the prosthetic leg to become the supporting leg again for lowering the body down for the next step with the prosthetic limb (5, 6). The side-step pattern is similar to the step-to pattern, but in addition the patient turns the whole body to one side to descend the slope not with a straight but oblique step-to pattern to further reduce the downhill-slope force to be controlled (3, 4, 6). Both patterns allow for only slow slope and hill descent, with the side-step pattern being even slower than the straight step-to pattern and expose the patient as a disabled person to the public. In subjects with MFCL-2 mobility, MPK have been shown to significantly reduce falls by up to 80% ($p < .05$ to $.01$) (1, 3, 8, 9), significantly improve validated indicators of the risk of falling such as the timed up and go test ($p = .018$ / $p = .001$) (5, 11), and balance confidence as assessed with the Activity-specific balance confidence (ABC) scale ($p = .001$) (5). Consequently, patients with MFCL-2 mobility have been demonstrated to significantly improve their downhill gait pattern and significantly increase their downhill walking speed by 27-36% ($p = .002$ / $< .001$) when using a MPK as compared to a non-MPK (1, 3, 5).

Evidence Summary

Kenevo[®] Microprocessor Stance Control Knee

References

1. Kannenberg A, Zacharias B, Pröbsting E: Benefits of microprocessor prosthetic knees to limited community ambulators: A systematic review. *J Rehabil Res Dev* 2014; 51 (10): 1469-1495.
2. Highsmith MJ, Kahle JT, Miro RM, Mengelkoch, MJ: Ramp descent performance with the C-leg and interrater reliability of the Hill Assessment Index. *Prosthet Orthot Int* 2013; 37(5): 362-368.
3. Hafner BJ, Smith DG: Differences in function and safety between Medicare Functional Classification Level-2 and -3 transfemoral amputees and influence of prosthetic knee joint control. *J Rehabil Res Dev* 2009; 46 (3): 417-434.
4. Hafner BJ, Willingham LL, Buell NC, Allyn KJ, Smith DG: Evaluation of Function, Performance, and Preference as Transfemoral Amputees Transition from Mechanical to Microprocessor Control of the Prosthetic Knee. *Arch Phys Med Rehabil* 2007; 88(2): 207-17.
5. Burnfield JM, Eberly VJ, Gronely JK, Perry J, Yule WJ, Mulroy SJ. Impact of stance phase microprocessor-controlled knee prosthesis on ramp negotiation and community walking function in K2 level transfemoral amputees. *Prosthet Orthot Int* 2012, 36 (1): 95-104.
6. Vrieling AH, Van Keeken HG, Schoppen T, et al. Uphill and downhill walking in unilateral lower limb amputees. *Gait Posture* 2008; 28(2): 235–242.
7. Blumentritt S: Biomechanical aspects of the indications of prosthetic knee joints [Biomechanische Aspekte zur Indikation von Prothesenkniegelenken]. *Orthopädie-Technik* 2004; 55(6): 508-524. Publication in German.
8. Kahle JT, Highsmith MJ, Hubbard SL: Comparison of Non-microprocessor Knee Mechanism versus C-Leg on Prosthesis Evaluation Questionnaire, Stumbles, Falls, Walking Tests, Stair Descent, and Knee Preference; *J Rehabil Res Dev* 2008; 45 (1): 1-14.
9. Kaufman KR, Bernhardt KA, Symms K. Functional assessment and satisfaction of transfemoral amputees with mobility (FASTK2): A clinical trial of microprocessor-controlled vs. non-microprocessor-controlled knees. *Clin Biomech (Bristol, Avon)* 2018 Oct;58:116-122. doi: 10.1016/j.clinbiomech.2018.07.012. Epub 2018 Jul 19.
10. Lansade C, Vicaut E, Paysant J, Ménager D, Cristina MC, Braatz F, Domayer S, Pérennou D, Chiesa G. Mobility and safety with a microprocessor-controlled knee in moderately active amputees: A multi-centric randomized crossover trial. *Ann Phys Rehabil Med*, 2018 May 14. pii: S1877-0657(18) 30058-7. doi: 10.1016/ j.rehab.2018.04.003. [Epub ahead of print]

Improved negotiation of uneven terrain and obstacles in the walkway

Negotiation of uneven terrain and clearance of obstacles in the walkway are common activities in daily living. As most non-microprocessor controlled knee mechanisms have been designed for ambulation on level ground (7-9), uneven terrain and obstacles in the walkway expose above-knee amputees to an increased risk of stumbling and falling (8-9). Therefore, many patients usually avoid walking on uneven terrain or walkways with obstacles, or negotiate them very cautiously and slowly. In subjects with MFCL-2 mobility, MPK have been shown to significantly reduce falls by up to 80% ($p < .05$ to $.01$) (1-4), significantly improve validated indicators of the risk of falling such as the timed up and go test ($p = .018$ / $p = .001$) (5, 6), and balance confidence as assessed with the Activity-specific balance confidence (ABC) scale ($p = .001$) (5). Consequently, timed walk tests on uneven terrain and obstacle courses have shown that patients with MFCL-2 mobility using MPK are able to negotiate these terrains at significantly faster walking speeds (2, 3). Uneven terrain may be negotiated 20% faster ($p = .001$) (3) and obstacle courses 11% faster ($p = .02$) without and 12 faster ($p = .02$) with a concurrent mental task (2). Thus, above-knee amputees with MFCL-2 mobility are able to negotiate uneven terrain and clear obstacles in the walkway significantly better and faster with a MPK than with any non-microprocessor controlled knee.

Evidence Summary

Kenevo® Microprocessor Stance Control Knee

References

1. Kannenberg A, Zacharias B, Pröbsting E: Benefits of microprocessor prosthetic knees to limited community ambulators: A systematic review. *J Rehabil Res Dev* 2014; 51 (10): 1469-1495.
2. Hafner BJ, Smith DG: Differences in function and safety between Medicare Functional Classification Level-2 and -3 transfemoral amputees and influence of prosthetic knee joint control. *J Rehabil Res Dev* 2009; 46 (3): 417-434.
3. Kahle JT, Highsmith MJ, Hubbard SL: Comparison of Non-microprocessor Knee Mechanism versus C-Leg on Prosthesis Evaluation Questionnaire, Stumbles, Falls, Walking Tests, Stair Descent, and Knee Preference; *J Rehabil Res Dev* 2008; 45 (1): 1-14.
4. Kaufman KR, Bernhardt KA, Symms K. Functional assessment and satisfaction of transfemoral amputees with mobility (FASTK2): A clinical trial of microprocessor-controlled vs. non-microprocessor-controlled knees. *Clin Biomech (Bristol, Avon)* 2018 Oct;58:116-122. doi: 10.1016/j.clinbiomech.2018.07.012. Epub 2018 Jul 19.
5. Burnfield JM, Eberly VJ, Gronely JK, Perry J, Yule WJ, Mulroy SJ. Impact of stance phase microprocessor-controlled knee prosthesis on ramp negotiation and community walking function in K2 level transfemoral amputees. *Prosthet Orthot Int* 2012, 36 (1): 95-104.
6. Lansade C, Vicaut E, Paysant J, Ménager D, Cristina MC, Braatz F, Domayer S, Pérennou D, Chiesa G. Mobility and safety with a microprocessor-controlled knee in moderately active amputees: A multi-centric randomized crossover trial. *Ann Phys Rehabil Med*, 2018 May 14. pii: S1877-0657(18) 30058-7. doi: 10.1016/j.rehab.2018.04.003. [Epub ahead of print]
7. Highsmith MJ, Kahle JT, Miro RM, Mengelkoch, MJ: Ramp descent performance with the C-leg and interrater reliability of the Hill Assessment Index. *Prosthet Orthot Int* 2013; 37(5): 362-368.
8. Blumentritt S, Schmalz T, Jarasch R: The safety of C-leg: Biomechanical tests. *J Prosthet Orthot* 2009, 21(1): 2-17.
9. Blumentritt S: Biomechanical aspects of the indications of prosthetic knee joints [Biomechanische Aspekte zur Indikation von Prothesenkniegelenken]. *Orthopädie-Technik* 2004; 55(6): 508-524. Publication in German.

Improved stair descent

Stairs are often encountered barriers in daily living and require greater lower-extremity range of motion and strength to negotiate, compared to level ground walking. Able-bodied people use reciprocal (step-over-step) stair descent, in which the supporting leg lowers the whole body down to the next step where the swinging leg becomes the supporting leg after landing. In above-knee amputees, most non-microprocessor controlled prosthetic knee mechanisms do not allow for any or enough knee flexion during weight bearing to lower the body with the supporting prosthetic leg or are too difficult to control for most patients to do so safely (4-6). That’s why above-knee amputees usually use a step-to pattern to descend stairs: The supporting sound leg lowers the body down using an eccentric contraction of the quadriceps to control knee flexion allowing the patient to land on the next step with the prosthetic leg. Then the sound leg is positioned on the same step next to the prosthetic leg to become the supporting leg again for lowering the body down to the next step (2, 4, 5, 6). This step-to pattern allows for only slow stair descent and exposes the patient as a disabled person. In subjects with MFCL-2 mobility, MPK have been shown to significantly reduce falls by up to 80% ($p < .05$ to $.01$) (1-3, 8), significantly improve validated indicators of the risk of falling such as the timed up and go test ($p = .018$ / $p = .001$) (7, 9), and balance confidence as assessed with the Activity-specific balance confidence (ABC) scale ($p = .001$) (7). Consequently, patients with MFCL-2 mobility have been demonstrated in several studies to significantly improve their gait quality when descending stairs ($p = .04$ / $p = .008$) and adopt a significantly more natural stair descent pattern in which the supporting prosthetic leg may even be used to lower the body down

Evidence Summary

Kenevo[®] Microprocessor Stance Control Knee

to the next step (1-3). This gait pattern is considerably faster than a step-to pattern and does not expose them as disabled persons to the public.

References

1. Kannenberg A, Zacharias B, Pröbsting E: Benefits of microprocessor prosthetic knees to limited community ambulators: A systematic review. *J Rehabil Res Dev* 2014; 51 (10): 1469-1495.
2. Hafner BJ, Smith DG: Differences in function and safety between Medicare Functional Classification Level-2 and -3 transfemoral amputees and influence of prosthetic knee joint control. *J Rehabil Res Dev* 2009; 46 (3): 417-434.
3. Kahle JT, Highsmith MJ, Hubbard SL: Comparison of Non-microprocessor Knee Mechanism versus C-Leg on Prosthesis Evaluation Questionnaire, Stumbles, Falls, Walking Tests, Stair Descent, and Knee Preference; *J Rehabil Res Dev* 2008; 45 (1): 1-14.
4. Hafner BJ, Willingham LL, Buell NC, Allyn KJ, Smith DG: Evaluation of Function, Performance, and Preference as Transfemoral Amputees Transition from Mechanical to Microprocessor Control of the Prosthetic Knee. *Arch Phys Med Rehabil* 2007; 88(2): 207-17.
5. Schmalz T, Blumentritt S, Marx B: Biomechanical Analysis of Stair Ambulation in Lower Limb Amputees. *Gait Posture* 2007; 25: 267-278.
6. Blumentritt S: Biomechanical aspects of the indications of prosthetic knee joints [Biomechanische Aspekte zur Indikation von Prothesenkniegelenken]. *Orthopädie-Technik* 2004; 55(6): 508-524. Publication in German.
7. Burnfield JM, Eberly VJ, Gronely JK, Perry J, Yule WJ, Mulroy SJ. Impact of stance phase microprocessor-controlled knee prosthesis on ramp negotiation and community walking function in K2 level transfemoral amputees. *Prosthet Orthot Int* 2012, 36 (1): 95-104.
8. Kaufman KR, Bernhardt KA, Symms K. Functional assessment and satisfaction of transfemoral amputees with mobility (FASTK2): A clinical trial of microprocessor-controlled vs. non-microprocessor-controlled knees. *Clin Biomech (Bristol, Avon)* 2018 Oct;58:116-122. doi: 10.1016/j.clinbiomech.2018.07.012. Epub 2018 Jul 19.
9. Lansade C, Vicaut E, Paysant J, Ménager D, Cristina MC, Braatz F, Domayer S, Pérennou D, Chiesa G. Mobility and safety with a microprocessor-controlled knee in moderately active amputees: A multi-centric randomized crossover trial. *Ann Phys Rehabil Med*, 2018 May 14. pii: S1877-0657(18) 30058-7. doi: 10.1016/j.rehab.2018.04.003. [Epub ahead of print]

Reduced cognitive demand / Improved multitasking capacity while walking

The need to execute a concurrent task while walking is a common activity in daily living. As most non-microprocessor controlled knee (NMPK) mechanisms have been designed for ambulation on level ground and require a permanent alertness of the patient to actively stabilize the knee (5-7), above-knee amputees usually spend a lot of concentration and mental energy on screening their walkway for any kind of perturbation (2, 5, 6, 7). Therefore, their capacity to execute a concurrent task while walking with the prosthesis is considerably limited. In subjects with MFCL-2 mobility, MPK have been shown to significantly reduce falls by up to 80% ($p < .05$ to $.01$) (2, 3, 8, 9), significantly improve validated indicators of the risk of falling such as the timed up and go test ($p = .018$ / $p = .001$) (4, 10), and balance confidence as assessed with the Activity-specific balance confidence (ABC) scale ($p = .001$) (4). Consequently, tests assessing the self-reported ability for multitasking while walking with the prosthesis demonstrated a significant improvement by 21% ($p = .04$) when patients with MFCL-2 mobility were using the C-Leg as compared to NMPK (2). An observational study with 29 subjects assessing Kenevo as compared to NMPK typically prescribed and fit in patients with MFCL-2 mobility found that 79% of participants reported much less or less concentration required to walk with Kenevo (1). Thus, MPK

Evidence Summary

Kenevo[®] Microprocessor Stance Control Knee

may reduce cognitive demand and improve capacity for multitasking while walking with the prosthesis in patients with MFCL-2 mobility.

References

1. Mileusnic M, Hahn A, Reiter S. Effects of a novel microprocessor-controlled knee, Kenevo, on the safety, mobility, and satisfaction of lower-activity patients with transfemoral amputation. *J Prosthet Orthot* 2017; 29(4): 198-205.
2. Hafner BJ, Smith DG: Differences in function and safety between Medicare Functional Classification Level-2 and -3 transfemoral amputees and influence of prosthetic knee joint control. *J Rehabil Res Dev* 2009; 46 (3): 417-434.
3. Kahle JT, Highsmith MJ, Hubbard SL: Comparison of Non-microprocessor Knee Mechanism versus C-Leg on Prosthesis Evaluation Questionnaire, Stumbles, Falls, Walking Tests, Stair Descent, and Knee Preference; *J Rehabil Res Dev* 2008; 45 (1): 1-14.
4. Burnfield JM, Eberly VJ, Gronely JK, Perry J, Yule WJ, Mulroy SJ. Impact of stance phase microprocessor-controlled knee prosthesis on ramp negotiation and community walking function in K2 level transfemoral amputees. *Prosthet Orthot Int* 2012, 36 (1): 95-104.
5. Highsmith MJ, Kahle JT, Miro RM, Mengelkoch, MJ: Ramp descent performance with the C-leg and interrater reliability of the Hill Assessment Index. *Prosthet Orthot Int* 2013; 37(5): 362-368.
6. Blumentritt S, Schmalz T, Jarasch R: The safety of C-leg: Biomechanical tests. *J Prosthet Orthot* 2009, 21(1): 2-17.
7. Blumentritt S: Biomechanical aspects of the indications of prosthetic knee joints [Biomechanische Aspekte zur Indikation von Prothesenkniegelenken]. *Orthopädie-Technik* 2004; 55(6): 508-524. Publication in German.
8. Kannenberg A, Zacharias B, Pröbsting E: Benefits of microprocessor prosthetic knees to limited community ambulators: A systematic review. *J Rehabil Res Dev* 2014; 51 (10): 1469-1495.
9. Kaufman KR, Bernhardt KA, Symms K. Functional assessment and satisfaction of transfemoral amputees with mobility (FASTK2): A clinical trial of microprocessor-controlled vs. non-microprocessor-controlled knees. *Clin Biomech (Bristol, Avon)* 2018 Oct;58:116-122. doi: 10.1016/j.clinbiomech.2018.07.012. Epub 2018 Jul 19.
10. Lansade C, Vicaut E, Paysant J, Ménager D, Cristina MC, Braatz F, Domayer S, Pérennou D, Chiesa G. Mobility and safety with a microprocessor-controlled knee in moderately active amputees: A multi-centric randomized crossover trial. *Ann Phys Rehabil Med*, 2018 May 14. pii: S1877-0657(18) 30058-7. doi: 10.1016/j.rehab.2018.04.003. [Epub ahead of print]

Improved overall mobility, especially in K2 patients

The more proximal the amputation, the greater is the physical and functional impairment to the individual, including a decreased likelihood of regaining household or community ambulation and an increased risk of falling (13-15). In subjects with an above-knee amputation, the prosthetic knee is a very important component, tasked with restoring knee biomechanics while at the same time providing maximum stability and safety. Most non-microprocessor controlled knee mechanisms have been designed for ambulation on level ground and require a permanent alertness of the patient to actively stabilize the knee in case of any perturbations (8-10). In subjects with MFCL-2 mobility, MPK have been shown to significantly reduce falls by up to 80% ($p < .05$ to $.01$) (1-3, 16), significantly improve validated indicators of the risk of falling such as the timed up and go test ($p = .018$ / $p = .001$) (4, 17), and balance confidence as assessed with the Activity-specific balance confidence (ABC) scale ($p = .001$) (4). Consequently, many patients are able to improve their overall mobility when using the C-Leg. Two studies demonstrated that 44% (3) or 50% (2), respectively, of patients with MFCL-2 mobility increased their overall mobility level to MFCL-3. Performance-based outcome measures suggest that these patients may be able to walk about 14-25% faster ($p = .01$ to $.000$) on level ground (1, 3, 5), around 20% quicker

Evidence Summary

Kenevo[®] Microprocessor Stance Control Knee

($p=.008$) on uneven surfaces (1, 3), and descend a slope 30% faster ($p=.002$ to $.001$) when using the C-Leg (1, 2, 4). Furthermore, negotiation of stairs is significantly improved ($p=.04$ to $.008$) (1-3) and patients are enabled to perform many activities of community ambulation and in the house that are considered typical of MFCL-3 mobility (1, 6, 7). In addition, an observational study with 29 subjects assessing Kenevo as compared to NMPK typically prescribed and fit in patients with MFCL-2 mobility found that when using the NMPK, 87% of patients reported to use a wheelchair with their prosthesis. When using Kenevo, only 37% of subjects ($p=.0046$) still required an additional wheelchair (18). It is therefore no longer justified to generally withhold microprocessor-controlled prosthetic knees from patients with MFCL-2 mobility.

References

1. Kannenberg A, Zacharias B, Pröbsting E: Benefits of microprocessor prosthetic knees to limited community ambulators: A systematic review. *J Rehabil Res Dev* 2014; 51 (10): 1469-1495.
2. Hafner BJ, Smith DG: Differences in function and safety between Medicare Functional Classification Level-2 and -3 transfemoral amputees and influence of prosthetic knee joint control. *J Rehabil Res Dev* 2009; 46 (3): 417-434.
3. Kahle JT, Highsmith MJ, Hubbard SL: Comparison of Non-microprocessor Knee Mechanism versus C-Leg on Prosthesis Evaluation Questionnaire, Stumbles, Falls, Walking Tests, Stair Descent, and Knee Preference; *J Rehabil Res Dev* 2008; 45 (1): 1-14.
4. Burnfield JM, Eberly VJ, Gronely JK, Perry J, Yule WJ, Mulroy SJ. Impact of stance phase microprocessor-controlled knee prosthesis on ramp negotiation and community walking function in K2 level transfemoral amputees. *Prosthet Orthot Int* 2012; 36 (1): 95-104.
5. Eberly VJ, Mulroy SJ, Gronley JK, Perry J, Burnfield JM. Impact of a stance phase microprocessor-controlled knee prosthesis on level walking in lower functioning individuals with transfemoral amputation. *Prosth Orthot Int* 2014; 38(6): 447-455.
6. Theeven PJ, Hemmen B, Geers RP, Smeets RJ, Brink PR, Seelen HA. Influence of advanced prosthetic knee joints on perceived performance and everyday life activity of low-functional persons with a transfemoral amputation or knee disarticulation. *J Rehabil Med* 2012; 44 (5): 454-461.
7. Theeven P, Hemmen B, Rings F, Meys G, Brink P, Smeets R, Seelen H. Functional added value of microprocessor-controlled knee joints in daily life performance of Medicare Functional Classification Level-2 amputees. *J Rehabil Med* 2011; 43 (10): 906-915.
8. Highsmith MJ, Kahle JT, Miro RM, Mengelkoch, MJ: Ramp descent performance with the C-leg and interrater reliability of the Hill Assessment Index. *Prosthet Orthot Int* 2013; 37(5): 362-368.
9. Blumentritt S, Schmalz T, Jarasch R: The safety of C-leg: Biomechanical tests. *J Prosthet Orthot* 2009, 21(1): 2-17.
10. Blumentritt S: Biomechanical aspects of the indications of prosthetic knee joints [Biomechanische Aspekte zur Indikation von Prothesenkniegelenken]. *Orthopädie-Technik* 2004; 55(6): 508-524. Publication in German.
11. Kaufman KR, Levine JA, Brey RH, et al. Gait and Balance of transfemoral amputees using passive mechanical and microprocessor-controlled prosthetic knees. *Gait Posture* 2007; 26: 489-493.
12. Davies B, Datta D. Mobility outcome following unilateral lower limb amputation. *Prosthet Orthot Int*. 2003; 27(3): 186–190.
13. Robbins CB, Vreeman DJ, Sothmann MS, Wilson SL, Oldridge NB. A review of the long-term health outcomes associated with war-related amputation. *Mil Med*. 2009; 174(6): 588–592.
14. Miller WC, Deathe AB, Speechley M, Koval J. The influence of falling, fear of falling, and balance confidence on prosthetic mobility and social activity among individuals with a lower extremity amputation. *Arch Phys Med Rehabil*. 2001; 82(9):1238–1244.
15. van Velzen JM, van Bennekom CA, Polomski W, Slootman JR, van der Woude LH, Houdijk H. Physical capacity and walking ability after lower limb amputation: A systematic review. *Clin Rehabil*. 2006; 20(11): 999–1016.

Evidence Summary

Kenevo[®] Microprocessor Stance Control Knee

16. Kaufman KR, Bernhardt KA, Symms K. Functional assessment and satisfaction of transfemoral amputees with mobility (FASTK2): A clinical trial of microprocessor-controlled vs. non-microprocessor-controlled knees. Clin Biomech (Bristol, Avon) 2018 Oct;58:116-122. doi: 10.1016/j.clinbiomech.2018.07.012. Epub 2018 Jul 19.
17. Lansade C, Vicaut E, Paysant J, Ménager D, Cristina MC, Braatz F, Domayer S, Pérennou D, Chiesa G. Mobility and safety with a microprocessor-controlled knee in moderately active amputees: A multi-centric randomized crossover trial. Ann Phys Rehabil Med, 2018 May 14. pii: S1877-0657(18) 30058-7. doi: 10.1016/j.rehab.2018.04.003. [Epub ahead of print]
18. Mileusnic M, Hahn A, Reiter S. Effects of a novel microprocessor-controlled knee, Kenevo, on the safety, mobility, and satisfaction of lower-activity patients with transfemoral amputation. J Prosthet Orthot 2017; 29(4): 198-205.

Reimbursement, Ottobock North America
P 800 328 4058 F 800 230 3962
professionals.ottobockus.com
professionals.ottobock.ca
reimbursement911@ottobock.com